

Listing#: E115838
Report #: 127053
Original Certification Date: January 31, 2024
Revised Certification Date:

This Certification is issued to:
Keit Limited
Unit 4, Zephyr Building, Eighth St, Harwell Campus,
Didcot OX11 0RL, United Kingdom

Stating that the product(s):
FTIR Spectrometer,
IRmadillo

Product Rating(s):
100-240 Vac, 50/60Hz, 120W

Achieved Certification to the following standard(s):
UL 60079-0 (Edition 7) 2019 rev 2020
UL 60079-1 (Edition 7) 2015
UL 60079-26 (Edition 3) 2017 Rev 2022
UL 61010-1 (Edition 3rd) 2018 Rev 2022
CSA C22.2 60079-0 (Edition 4) 2019
CSA C22.2 60079-1:2016
CSA C22.2 60079-26: 2022
CSA C22.2 No. 61010-1-12/A1 (Edition 3) 2018 rev 2022



Jeremy Maxwell
Eurofins Electrical and Electronic Testing North America, Inc.

All changes proposed in the previously identified product that affects the above information must be submitted to Eurofins for evaluation prior to implementation to assure continued NRTL Certification status. The covered product(s) shall be subject to follow-up inspections to ensure that the Certified product(s) are identical to the product sample evaluated by Eurofins E&E NA and that all responsibilities are being fulfilled as specified in the Applicants' Responsibility section of the Certification Report. The Applicant named above has been authorized Eurofins E&E NA to represent the product(s) listed in this record as "MET Certified" and to mark this/these product(s) according to the terms and conditions of the Eurofins E&E NA Applicant Contract, Listing Reports, and the applicable agreements. Only the product(s) bearing the MET Mark and under a follow-up service are considered to be included in this Certification program. This certification has been granted under a System 3 program as defined in ISO/IEC 17067.



Eurofins E&E North America, Inc. is accredited by OSHA and the Standards Council of Canada.



N
R
T
L